## **Uniphos Detector Tube Data Sheet**Direct Reading Length of Stain Type

SAM-9 **Ammonia**  $NH_3$ 

TWA (TLV): 25 ppm STEL (TLV): 35 ppm Flammable Range: 15-28%

## Specification:

Measurement Range	Extended	Standard	Extended
	0.5 – 15 %	1 - 30 %	2 -60 %
No. of Pump Strokes	2 (200 mL)	1 (100 mL)	0.5 (50 mL)
Volume Correction Factor (VCF)*	0.5	1.0	2.0
Sampling Time	1 minute per pump stroke (100 mL)		
Color Change	Yellow → Violet		
Shelf Life	2 years		
Storage condition	Below 10°C (50°F)		
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.		
Reaction Principle	NH <sub>3</sub> + Acid → Reaction product		
Relative standard deviation	± 10 %	•	

<sup>\*</sup> Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

## Possible Interferences:

Compound	Concentration (%)	Interference	Colour change / Comments
Amines		+	Yellow to Blue
Hydrazine		+	Yellow to Blue
CO <sub>2</sub>	10	No	No Effect

## **Correction for Environmental Parameters:**

Temperature	Not necessary between 0 - 40°C (32 - 104°F).	
Relative humidity	Not necessary between 10 - 90 %.	