

# Uniphos Detector Tube Data Sheet

Direct Reading Length of Stain Type

## Carbon Dioxide $\text{CO}_2$

## SCD-7

TWA (TLV): 5000 ppm

STEL (TLV): 30,000 ppm

Non Flammable

### Specification:

	Extended	Standard	Extended
Measurement Range	0.125-1.5%	0.25-3%	0.5-6%
No. of Pump Strokes	2 (200 mL)	1 (100 mL)	0.5 (50 mL)
Volume Correction Factor (VCF)*	0.5	1.0	2.0
Sampling Time	2.5 minutes per pump stroke (100 mL)		
Color Change	Blue → Off white		
Shelf Life	2 years		
Storage condition	Below 10°C (50°F)		
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.		
Reaction Principle	$\text{CO}_2 + \text{NaOH} \rightarrow \text{NaCO}_3 + \text{H}_2\text{O}$		
Relative standard deviation	± 10 %		

\* Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

### Possible interferences:

Compound	Concentration (%)	Interference	Colour change / Comments
CO	1.5	No	No Effect
H <sub>2</sub> S	0.2	+	
NH <sub>3</sub>	5	No	No Effect
Benzene	0.01	No	Dark blue ring
SO <sub>2</sub>	0.02	+	

### Correction for Environmental Parameters:

Temperature (°C) :	4	20	30	40
Temperature Correction Factor (TCF) *	0.82	1	1.03	1.13
Relative humidity	Not necessary between 10 - 90 %.			

\* Multiply the observed reading by the correction factor (VCFXTCF) to obtain the true concentration.



**Manufactured By:** Uniphos Envirotronic Pvt. Ltd.  
Nahuli (PO), Umbergaon (Tal.)  
Valsad (Dist), Gujarat, India – 396108

Rev.0812