

Uniphos Detector Tube Data Sheet

Direct Reading Length of Stain Type

Mercury Vapour Hg

TWA (TLV): 0.025 mg/m³

STEL (TLV): NA

SHG-1L

Flammable Range: NA

Specification:

Measurement Range	Extended	Standard	Extended
	Do not extend	0.1- 2 mg/m ³	0.35 – 7 mg/m ³
No. of Pump Strokes		5 (500 mL)	1 (100 mL)
Volume Correction Factor (VCF)*		1	3.5
Sampling Time	2.0 minutes per pump stroke (100 mL)		
Color Change	White → Pale orange		
Shelf Life	1 year		
Storage condition	Below 10°C (50°F)		
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.		
Reaction Principle	$\text{Hg} + 2\text{Cu}_2\text{I}_2 \rightarrow \text{Cu}_2(\text{HgI}_4) + 2\text{Cu}$		
Relative standard deviation	± 10 %		

* Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

Possible Interferences:

Compound	Concentration (ppm)	Interference	Colour change / Comments
H ₂ S		+	White to Pale orange
CO ₂	5000	No	No Effect

Correction for Environmental Parameters:

Temperature	Not necessary between 0 - 40°C (32 - 104°F).
Relative humidity	Not necessary between 10 - 90 %.



Manufactured By: Uniphos Envirotronic Pvt. Ltd.
Nahuli (PO), Umbergaon (Tal.)
Valsad (Dist), Gujarat, India – 396108

Rev.0812