

# Uniphos Detector Tube Data Sheet

Direct Reading Length of Stain Type

## Oxygen

TWA (TLV): NA

## O<sub>2</sub>

STEL (TLV): NA

## SOX-9M

Flammable Range: NA

### Specification:

	Extended	Standard	Extended
Measurement Range	Do not extend	3-24%	Do not extend
No. of Pump Strokes		0.5 (50 mL)	
Volume Correction Factor (VCF)*		1.0	
Sampling Time	2 minutes per ½ pump stroke (50 mL)		
Color Change	Dark violet → White		
Shelf Life	2 years		
Storage condition	Below 10°C (50°F)		
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.		
Reaction Principle	$O_2 + 4TiCl_3 + 6H_2O \rightarrow 4TiO_2 + 12 HCl$		
Relative standard deviation	± 10 %		

\* Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

### Possible Interferences:

Compound	Concentration (ppm)	Interference	Colour change / Comments
Carbon dioxide	20	No	No Effect
Hydrogen sulphide	4.7	No	No Effect

### Correction for Environmental Parameters:

Temperature	Not necessary between 0 - 40°C (32 - 104°F).
Relative humidity	Not necessary between 10 - 90 %.



**Manufactured By:** Uniphos Envirotronic Pvt. Ltd.  
Nahuli (PO), Umbergaon (Tal.)  
Valsad (Dist), Gujarat, India – 396108

Rev.0812