

Phosphoric Acid Esters 0.05/a

Catalog No 6727461

Standard Measuring Range: 0.05 ppm Dichlorvos
Number of Strokes (n): 10
Time for Measurement: about 5 minutes
Standard Deviation: ± 30%
Color Change: yellow to red

Ambient Operation Conditions

Temperature: 0 to 40°C
Absolute Humidity: 3 to 18 mg H₂O/L

Reaction Principle

- $(\text{CH}_3\text{O})_2\text{PO}_2\text{-CH=CCl}_2 + \text{Cholinesterase} \longrightarrow \text{inactive enzyme}$
- $\text{Butyrylcholine iodide} + \text{H}_2\text{O} \longrightarrow \text{Butyric acid}$
- $\text{Butyric acid} + \text{Phenol red} \longrightarrow \text{yellow reaction product}$

If phosphoric acid esters are present the enzyme is inactivated and butyric acid will not form, thus the weak alkali solution colors the indicating layer red and must be stable for 1 minute.

If the enzyme remains active, phosphoric acid esters are not present and the indicating layer remains yellow because of butyric acid formation.

Cross Sensitivity

Other phosphoric acid esters than dichlorvos are also indicated, but with different sensitivities.

Additional Information

After performing the required 10 pump strokes the reagent ampule must be broken and the liquid transferred to the enzyme layer by gently tapping the side of the tube. The substrate layer must not become wetted! After waiting one (1) minute the liquid must be carefully drawn up the marking line using the pump. Wait another minute before drawing the liquid onto the indication layer using the pump.



Dräger

Manufactured by Draeger Safety

Distributed by AFC International, Inc.

PO Box 894

DeMotte, IN 46310

800.952.3293 • 219.987.6825 • fax 219.987.6826

www.afcintl.com • sales@afcintl.com