

WARRANTY EXCLUSIONS AND LIMITATION OF LIABILITY

IF THE USER DOES NOT ACCEPT THE FOLLOWING TERMS, THE USER SHOULD NOT USE THE SAFEAIR BADGES.

Seller warrants that, for the lesser of one (1) year or the shelf-life provided on the Seller's product, its packaging or its literature, its products shall conform to the description of such products as provided on the Seller's product, its packaging or its literature. THIS WARRANTY IS EXCLUSIVE, AND SELLER MAKES NO OTHER WARRANTY, EXPRESS OR IMPLIED, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR ANY PARTICULAR PURPOSE. Seller's warranties made in connection with this sale shall not be effective if Seller has determined, in its sole discretion, that Buyer has misused the products in any manner, has failed to use the products in accordance with safe industry standards and practices, or has failed to use the products in accordance with Seller's written operating instructions, or other instructions furnished by Seller to Buyer.

Seller's sole and exclusive liability and Buyer's exclusive remedy with respect to products to Seller's satisfaction to be defective or nonconforming shall be replacement of such products without charge or refund of the purchase price, in Seller's sole discretion, upon the return of such products in accordance with Seller's instructions. All claims must be brought within one (1) year of shipment, regardless of their nature.

SELLER SHALL NOT IN ANY EVENT BE LIABLE FOR INCIDENTAL, CONSEQUENTIAL OR SPECIAL DAMAGES OF ANY KIND RESULTING FROM ANY USE OR FAILURE OF THE PRODUCTS, EVEN IF SELLER HAS BEEN ADVISED OF THE POSSIBILITY OF SUCH DAMAGE INCLUDING, WITHOUT LIMITATION, LIABILITY FOR LOSS OF USE, LOSS OF WORK IN PROGRESS, DOWN TIME, LOSS OF REVENUE OR PROFITS, FAILURE TO REALIZE SAVINGS, LOSS OF PRODUCTS OF BUYER OR OTHER USE OF ANY LIABILITY OF BUYER TO A THIRD PARTY ON ACCOUNT OF SUCH LOSS, OR FOR ANY LABOR OR ANY OTHER EXPENSE, DAMAGE OR LOSS OCCASIONED BY SUCH PRODUCT INCLUDING PERSONAL INJURY OR PROPERTY DAMAGE UNLESS SUCH PERSONAL INJURY OR PROPERTY DAMAGE IS CAUSED BY SELLER'S GROSS NEGLIGENCE.

Warning: Keep out of reach of children, if ingested seek medical attention immediately.



2557 Production Rd.
Virginia Beach, VA 23454
Phone: (757) 431-2260 Fax: (757) 216-6209
e-mail: customerservice@morphtec.com

©1997, Morphix Technologies®
All rights reserved
Publication Number: KM1-20A.4-0714
US Patent Number: 6,284,198



Operating Instructions For Ozone Monitor (Part Number: 382004)

Technical Summary

Physical Specifications:

Dimensions	(74 ± 1 mm) x (41 ± 1 mm) x (1 mm)
Weight	1.5 g
Refrigerated shelf life	1 year
Color change	blue to white

Sampling Parameters:

Exposure level	0.05 ppm-hr
Minimum detectable limit (8 hours)	0.006 ppm
Maximum recommended sampling time	48 hours
Minimum recommended sampling time	15 minutes
Relative humidity range	30% - 80%
Face velocity range	10 – 150 cm/sec
Temperature range	60°F - 91°F
Light effect – UV (direct sunlight)	not recommended
Light effect – visible	no effect
Color stability	4 weeks

Applications:

The SafeAir ozone badge may be used for personnel or area monitoring for exposure times ranging from 15 minutes to 48 hours.

Cross Interferences:

Chlorine does not affect the performance of the monitor. Hydrogen peroxide is a known interference. Up to 0.3 ppm nitrogen dioxide shows no interference. Exposure to 0.5 ppm nitrogen dioxide for 5 hours causes false positive readings equivalent to 0.04 ppm ozone; exposure to 1 ppm nitrogen dioxide for 3 hours causes false positive readings equivalent to 0.04 ppm of ozone. No further interferences are known.

Introduction

Ozone is a bluish gas with a pleasant, characteristic odor in concentrations of less than 2 ppm. In higher concentrations, the odor is pungent and irritating. Ozone is a severe irritant to the eyes and the mucous membranes. Long-term exposures will cause pulmonary edema (abnormal fluid build up in the lungs) and chronic respiratory disease. OSHA and NIOSH exposure limit for ozone is 0.1 ppm (TWA).

Ozone is used as a disinfectant for air and water. It is also used to bleach waxes, textiles and oils.

Principle of Operation

The SafeAir ozone badge is a monitoring system designed to indicate the presence of ozone at concentrations below the permissible exposure limit. The SafeAir ozone badge detects the presence of ozone by forming a color change in the shape of an exclamation mark inside the triangle. This indication is produced by a color-forming reaction which occurs when ozone reacts with a flat indicator layer.

Operating Instructions

1. Remove the pouch from the refrigerator and allow it to warm to room temperature.
2. Remove the badge from its protective pouch. Remove activation label before monitoring.
3. Remove activation label before monitoring.
4. For personnel monitoring, attach the badge near the user's breathing zone (i.e. the collar) with the front side exposed to the surrounding atmosphere.
5. For area monitoring, attach the badge to a stand and mount in a centralized area with the front side exposed to the surrounding atmosphere.
6. The exclamation mark appears within the triangle when ozone is present. Please note that the exclamation mark will appear underneath the printed exposure dose (sensitivity).
7. To obtain the average concentration, divide the exposure dose (ppm-hr) by the exposure time in hours (hr).

Storage

The SafeAir ozone badge should be refrigerated in its sealed bag at all times.

Benefits

1. Accurate Detection: The SafeAir ozone badge is designed to react selectively with ozone with minimum interference from other substances.
2. Applications: The SafeAir badge may be used for personnel screening and for area monitoring or area mapping.
3. Ease of Use: The SafeAir badge is a direct-read device that gives immediate, on-site results.

Other Available Monitors

1. SafeAir Badges:

Ammonia	Formaldehyde	Phosgene
Aromatic Isocyanates	Hydrazine	Phosgene Dura
Carbon Monoxide	Hydrogen Chloride	Phosgene Medi
Chlorine	Hydrogen Sulfide	Sulfur Dioxide
Chlorine/Chlorine Dioxide	Mercury	
2. SafeAir Color Comparators:

Chlorine	Hydrogen Chloride	Phosgene Dura
Chloroformates ¹	Phosgene	Phosgene Medi
Hydrazine	Phosgene ext. range	TDI ²

If you require SafeAir monitors for a chemical hazard not listed, please contact Morphix Technologies® for a free compound consultation at (800) 808-2234.

¹ To be used with the SafeAir phosgene badges

² To be used with the SafeAir aromatic isocyanates badges